

Compact Electron Paramagnetic Resonance on a Chip Spectrometer Using a Single Sided Permanent Magnet

Michele Segantini, Gianluca Marcozzi, Tarek Elrifai, Ekaterina Shabratova, Katja Höflich, Mihaela Deaconea, Volker Niemann, Rainer Pietig, Joseph E. McPeak,* Jens Anders, Boris Naydenov, and Klaus Lips



Cite This: *ACS Sens.* 2024, 9, 5099–5108



Read Online

ACCESS |




Metrics & More



Article Recommendations



Supporting Information



ABSTRACT: Electron paramagnetic resonance (EPR) spectroscopy provides information about the physical and chemical properties of materials by detecting paramagnetic states. Conventional EPR measurements are performed in high Q resonator using large electromagnets which limits the available space for operando experiments. Here we present a solution toward a portable EPR sensor based on the combination of the EPR-on-a-Chip (EPRoC) and a single-sided permanent magnet. This device can be placed directly into the sample environment (i.e., catalytic reaction vessels, ultrahigh vacuum deposition chambers, aqueous environments, etc.) to conduct in situ and operando measurements. The EPRoC reported herein is comprised of an array of 14 voltage-controlled oscillator (VCO) coils oscillating at 7 GHz. By using a single grain of crystalline BDPA, EPR measurements at different positions of the magnet with respect to the VCO array were performed. It was possible to create a 2D spatial map of a $1.5\text{ mm} \times 5\text{ mm}$ region of the magnetic field with $50\text{ }\mu\text{m}$ resolution. This allowed for the determination of the magnetic field intensity and homogeneity, which are found to be 254.69 mT and 700 ppm, respectively. The magnetic field was mapped also along the vertical direction using a thin film a-Si layer. The EPRoC and permanent magnet were combined to form a miniaturized EPR spectrometer to perform experiments on tempol (4-hydroxy-2,2,6,6-tetramethylpiperidin-1-oxyl) dissolved in an 80% glycerol and 20% water solution. It was possible to determine the molecular tumbling correlation time and to establish a calibration procedure to quantify the number of spins within the sample.

KEYWORDS: EPR, EPRoC, single side permanent magnet, spin counting, in situ, in vivo, operando, molecular tumbling

Electron paramagnetic resonance (EPR) is a spectroscopic technique that detects paramagnetic species and free radicals, allowing the quantification and identification of physical, chemical, and biological substances. EPR spectroscopy is highly sensitive, quantitative, noninvasive and enables the analysis of various samples without complex preparation. Among other applications, it is possible by means of EPR to probe the free radicals in the human body for quantification of reactive oxygen species,^{1–4} in edible oil to monitor oxidative stability,^{5–8} in beer to ensure flavor stability,^{9–12} to detect defects in semiconductors^{13–15} and to measure the concentration of paramagnetic centers in vanadium redox flow batteries for monitoring the state of charge and health of the

battery.^{16,17} Most commercially available EPR devices are bulky and power-hungry due to the presence of electromagnets that produce the external magnetic field B_0 . A typical EPR experiment requires the use of microwave sources and detectors in combination with high-quality factor (Q) microwave resonators (active volume $\sim 250\text{ mm}^3$ at 9 GHz),

Received: April 5, 2024

Revised: September 10, 2024

Accepted: September 24, 2024

Published: September 26, 2024






Figure 1. (a) The permanent magnet was assembled by gluing together small segments of samarium–cobalt, shown here with different colors highlighting their orientation, to generate a homogeneous magnetic field B_0 parallel to the surface of the magnet. The dimensions of the magnet are shown (40 mm \times 25 mm \times 8 mm). (b) A schematic of the aluminum cage used to provide structural support for the permanent magnet. On the top of the central region of the magnet, an etched μ -metal foil (orange) of approximately 1 mm \times 40 mm \times 0.05 mm is placed to improve the magnetic field homogeneity. The black pattern on top of the magnet corresponds to the etched region of the μ -metal. The final dimensions of the assembled magnet are shown (41 mm \times 35 mm \times 8 mm).

where the sample is placed inside. Due to the large Q -factor, microwave resonators typically have limited bandwidth, therefore the electromagnets are used to sweep the B_0 field, while keeping the microwave frequency constant to measure the EPR spectrum. Even though EPR has been proven to be a useful tool for some *in vivo*, *in situ* and *operando* experiments, for example to study pO_2 distribution,^{3,4} lithium-ion batteries,^{18,19} catalyst materials^{20–23} and for imaging,^{24–26} the implementation in conventional EPR spectrometers is challenging and limit these application. In most cases, samples must be properly designed to respect the geometrical constraints of the microwave resonator.^{18,20,27} Moreover, polar solvents must be avoided due to their large dielectric constant, which would lower the Q -factor of the resonator. A promising solution to overcome these limitations is provided by EPR-on-a-chip (EPRoC) technology, where the entire apparatus for EPR sensing is reduced to a small, portable, and versatile device that can be directly installed in the sample environment.^{28,29} The EPRoC is comprised of an array of voltage-controlled oscillator (VCO) coils that oscillate at a specific microwave frequency, allowing the replacement of the resonator with the VCO array for coupling of microwaves into the sample environment while maintaining a high apparent Q -factor.^{15,30,31} Moreover, due to the possibility of sweeping the frequency of the VCO array to perform EPR measurements,^{32,33} the portability of the EPRoC is further enhanced when used in combination with a permanent magnet.³⁴

In this work, a small single-sided permanent magnet has been developed with the aim to produce a static magnetic field B_0 of approximately 250 mT to satisfy the resonance condition for small organic radicals ($g \approx 2$) using a VCO array oscillation frequency of ~ 7 GHz. The goal of designing this permanent magnet was to minimize its size and weight to ensure high versatility and allow the installation together with the EPRoC in complex sample environments, such as in ultrahigh vacuum (UHV) evaporation chambers and catalytic reactors.³⁵ The VCO array produces the oscillating out-of-plane magnetic B_1 field and was designed employing a 0.13 μm BiCMOS technology with 14 coils.³⁶ The number of coils was extended compared to previous works³⁷ to improve the signal-to-noise ratio (SNR)³⁵ of the EPR measurements, achieving spin sensitivity of 2×10^{12} spins/ $\sqrt{\text{Hz}}$, as reported in ref 36. In this report, a single coil of the VCO array was used as a sensor to create a 2D image of the magnetic field intensity of the

permanent magnet. These results highlight the capabilities of the EPRoC to fully determine the intensity and the homogeneity of the magnetic field with high spatial resolution (50 μm), demonstrating a 10-fold improvement in the spatial resolution compared to previous reports.²⁶ In addition, the whole array of the EPRoC was used to map the magnetic field along the vertical direction. The EPRoC and the permanent magnet were then combined to measure tempol (4-hydroxy-2,2,6,6-tetramethylpiperidin-1-oxyl) dissolved in a mixture of glycerol and water with the aim to evaluate the possibility of performing quantitative measurements. As expected theoretically, the results of this investigation show a linear dependence of the signal intensity from the sample concentration, as confirmed by a commercial benchtop EPR system, allowing the calibration of the EPRoC for spin quantification. Finally, the tumbling correlation time of tempol in solution was determined with the EPRoC and the permanent magnet and compared to the commercial benchtop EPR spectrometer. This experiment was conducted to evaluate the capability of using the EPRoC for molecular tumbling analysis and viscosity determination.^{38,39}

Overall, these experiments using the EPRoC technology serve as a proof-of-concept to open new pathways to applications that are not accessible with conventional EPR spectrometers. Among these applications, the EPRoC can be used as an alternative solution to the commonly used magnetic field cameras based on magnetometry⁴⁰ and to design and test devices for magnetic cell manipulation.⁴¹ Moreover, EPRoC can be implemented directly into the sample environment and used as a “dipstick” sensor⁴² or as mobile surface explorer⁴³ to determine the number of spins and to measure the molecular tumbling rate of spin labels for *in vivo*, *in situ* and *operando* EPR experiments.

MATERIALS AND METHODS

Permanent Magnet. A small single-sided permanent magnet has been designed to be combined with the 7 GHz EPRoC array. The decision to design a single-sided permanent magnet for use with the EPRoC was driven primarily by the need for enhanced accessibility and integration for *in situ* experiments. This geometry allows for a more flexible access to the sample area, which is crucial for the integration of the EPRoC and permanent magnet in various experimental apparatuses, enhancing its applicability in a broad range of *in situ* studies where conventional multisided magnets would be too bulky or restrictive. The magnet geometry was determined to

respect the design criteria of a magnetic field intensity of 250 mT with a homogeneous region (≈ 100 ppm) of approximately $8\text{ mm} \times 1\text{ mm} \times 0.5\text{ mm}$ above the central area in the x , y and z directions, respectively. The magnetic materials (samarium–cobalt) were selected to ensure temperature stability of the magnetic field in the range of $20\text{--}150\text{ }^{\circ}\text{C}$ and compatibility with UHV to allow the installation in UHV deposition chambers. Simulations of the magnet geometry were performed by combining analytical solutions of the block-shaped magnet segments with uniform magnetization (Wolfram Mathematica) and Finite Element Analysis (FEM—Ansys Maxwell). Figure 1a shows a schematic of the permanent magnet, which was fabricated by gluing (Delo MONOPOX AD295) together 32 small segments (from 27 to 364 mm^3) of samarium–cobalt. Each segment was oriented in such a way as to generate a magnetic field B_0 parallel to the surface of the magnet. This orientation was required to fulfill the orthogonality condition between the B_0 field and the microwave magnetic field B_1 generated by the EPRoC to perform EPR measurements. The final dimensions of the assembled magnet are $40\text{ mm} \times 25\text{ mm} \times 8\text{ mm}$. The magnet was covered (Figure 1b) with an aluminum cage to provide additional support and to increase rigidity of the assembled structure to better resist mechanical stress. After the fabrication process, the magnetic field was measured using a Hall-probe sensor (3-axis Hall Probe C—Senis AG) and a Teslameter (3MH6 High-Precision, Low-Noise Teslameter—Senis AG) and the homogeneity estimated as the difference of the observed maximum and minimum values of magnetic field intensity, was found to be approximately 10,000 ppm. To further improve the homogeneity, a μ -metal strip was placed in the central region on the top of the magnet (Figure 1b) to locally weaken the magnetic field and thereby increase the overall homogeneity. The pattern of the μ -strip was etched using a laser (Laser StarFiber 150 P SM from COHERENT). The laser-etched pattern in the μ -metal foil allows for precise control of flux distribution enhancing the uniformity of the magnetic field and improving the overall field homogeneity in the region of interest. After placement of the μ -metal foil, the obtained homogeneity was approximately 1000 ppm. The UHV compatibility of the permanent magnet was examined by placing the permanent magnet in an UHV chamber, in which values of pressure below 7×10^{-9} mBar have been achieved.

Samples. Three different types of samples were used for the experiments. A single grain of α,γ -bis(diphenylene)- β -phenylallyl (BDPA complex, 1:1 with benzene, purchased from Sigma-Aldrich) was used as a point-like sample to perform the 2D mapping of the static magnetic field B_0 of the permanent magnet. The diameter of the sample ($\approx 200\text{ }\mu\text{m}$) was estimated from an image of the sample placed on a single coil recorded using an optical microscope (Motic, SMZ168 Series). A $15\text{ }\mu\text{m}$ thin film unhydrogenated amorphous silicon (a-Si) sample deposited using electron-beam evaporation (emission current: 560 mA, $T = 680\text{ }^{\circ}\text{C}$, deposition rate $\sim 400\text{--}450\text{ nm/min}$, sample rotation 15 rpm) from a solid source of undoped polycrystalline silicon on a $5\text{ cm} \times 5\text{ cm} \times 500\text{ }\mu\text{m}$ quartz substrate was used as a quasi-two-dimensional sample to cover the entire VCO array to determine the strength of the B_0 field while increasing the distance between the array and the permanent magnet. The paramagnetic silicon dangling bonds defects are homogeneous distributed within the $15\text{ }\mu\text{m}$ thin a-Si film. The sample was then cut into small pieces ($1\text{ mm} \times 2.5\text{ mm}$) in order to fit on the array sensor using a dicing saw (DISCO DAD3220). Four samples with different tempol (4-hydroxy-2,2,6,6-tetramethylpiperidin-1-oxyl) concentrations (63, 31.7, 15.8 and 7.9 mM) dissolved in a solution of 80% glycerol and 20% water were also used for the experiments. The tempol solutions were placed in thin flat capillaries ($50\text{ }\mu\text{m} \times 1\text{ mm} \times 2\text{ mm}$), which were sealed with a vinyl plastic compound (Critoseal). The wall thickness of the flat capillaries was $50\text{ }\mu\text{m}$.

Experimental Configuration. The EPRoC sensor used in this work was based on an LC VCO array designed to oscillate at 7 GHz and fabricated using a $0.13\text{ }\mu\text{m}$ BiCMOS technology.³⁶ Compared to previous works,³⁷ the number of coils was extended to 14, as the signal-to-noise ratio (SNR) increases $\propto \sqrt{N_{\text{coils}}}$, as reported in.^{35,37}

The estimated spin sensitivity of the array is $2 \times 10^{12} \frac{\text{spins}}{\sqrt{\text{Hz}}}$.³⁶ The EPRoC was bonded onto a single printed circuit board (PCB), which was then placed below the permanent magnet. The experimental configuration is comprised of the translation elements required to calibrate the position of the permanent magnet and the instruments necessary to perform EPR measurements using the EPRoC. In Figure 2a the block diagram of the experimental configuration is shown. The reference frequency f_{ref} for the EPR measurement was supplied to the EPRoC using a signal generator (Rohde & Schwarz SMB100B). The electrical power was supplied by a voltage source (Keysight E3646A) and the EPR signal was recorded using a lock-in amplifier by monitoring the variation of the V_{tune} (Anfatec eLockIn 203). The f_{ref} defines the oscillation frequency f_{VCO} of the VCO array according to the relation $f_{\text{VCO}} = 4 \times f_{\text{ref}}$ where the factor 4 is given by the frequency divider integrated into the EPRoC. The current I_{VCO} flowing through the VCO produces the microwave B_1 field, which is used to bidirectionally excite and detect the electron spins. The variation of the magnetization of the sample is detected as EPR signal by monitoring the variation of the oscillation frequency of the VCO. In this way, it is possible to sweep the VCO frequency to detect the EPR signal by recording the tuning voltage V_{tune} of the phase-locked loop (PLL). The VCO frequency can be sinusoidally modulated to detect the EPR signal

$$f_{\text{FM}}(t) = f_{\text{VCO}} + f_{\text{dev}} \cos(2\pi f_{\text{mod}} t) \quad (1)$$

where the f_{FM} is the modulated microwave frequency, f_{dev} is the frequency deviation, and f_{mod} is the modulation frequency, which is used as an external reference frequency for the lock-in amplifier. The frequency swept FM-detected EPR signal yields a dispersion-like derivative line shape.³⁴ Figure 2b shows a 3D model of the experimental realization of the scanning setup that was used to map the magnetic field homogeneity of the permanent magnet. The PCB, shown in green, has a T shape where the shorter side is 40 mm in length and the longer side is 78 mm in length. All electronic components necessary to operate the EPRoC, which is placed on the top side of the PCB and is shown in yellow, are placed on the back side of the PCB. This allowed the possibility of placing the permanent magnet in closer proximity to the EPRoC to perform the EPR measurements for imaging the magnetic field. The position of the magnet was varied by way of three DC servo motors (KDC101 ThorLabs) which are controlled with $1\text{ }\mu\text{m}$ precision over a maximum distance of 25 mm. The PCB was mounted on a long-traveling range translation stage (VAP10/M ThorLabs), which allowed for movement of the EPRoC along the vertical direction. The permanent magnet was fixed on a 3D printed mounting bracket fastened to three motorized translation stages (MT1/M-Z8 ThorLabs) to allow for movement of the magnet with respect to the EPRoC in three spatial directions (x , y , z). All necessary components were placed on an optical table to ensure mechanical stability during the measurements. To obtain reference spectra, frequency swept EPRoC measurements were performed by inserting the EPRoC and the PCB between the poles of an electromagnet a Bruker EPR300 X-band spectrometer.³² An enlarged view of the VCO array with 14 octagonal coils is shown in the inset. Depending on the experiment, the sample can be placed either on a single coil (orange circle) or onto the entire array. The PLL is embedded in the EPRoC, whose architecture is similar to the one reported in.³⁵

EPRoC Measurements. The EPR measurements of the three samples were performed using the EPRoC in the frequency sweep mode.⁴⁴ A single grain of BDPA was placed on a single coil of the VCO array to map the magnetic field of the permanent magnet. The sample was first measured in the electromagnet to determine the line width observed when the magnet homogeneity is $\approx 40\text{ ppm}$. As it has been reported in previous works, the EPRoC used in the frequency sweep mode has a baseline drift induced by the conversion of the tuning voltage to the oscillation frequency of the VCO.⁴⁴ The EPR signal is recorded by monitoring V_{tune} using a lock-in amplifier, and the baseline drift can be described by an arbitrary polynomial function (approximate fourth order). Therefore, to perform the baseline




Figure 2. (a) Block diagram of the EPRoC and necessary components. The RF generator supplies the reference frequency f_{ref} to the PLL integrated on the EPRoC. The current in the VCO array coils I_{VCO} generates the B_1 field. V_{tune} is the tuning voltage. When the sample goes into resonance, the oscillation frequency of the VCO changes along with the V_{tune} , which is measured by a lock-in amplifier to record the EPR signal. The modulation frequency f_{mod} is supplied from the RF generator to the lock-in amplifier. The permanent magnet (blue) is placed above the EPRoC. (b) 3D model of the scanning EPRoC setup with the permanent magnet. The PCB (green) with the VCO array (yellow) is mounted on a translation stage that can be moved along the vertical direction. The magnet is positioned above the PCB and is held by a 3D printed mounting bracket, which is connected to three DC servo motors which are used to control position in the x -, y -, and z -direction with respect to the EPRoC. In the rectangular dashed box, an enlarged view of the 14-coil EPRoC array is shown. The orange circle represents the BDPA grain that was placed on the coil for the magnetic field mapping. The PLL integrated on the EPRoC is indicated.

correction, an “on-resonance” spectrum was recorded by fixing the magnetic field intensity at $B_0 = 255.7$ mT. An “off-resonance” spectrum, measured at $B_0 = 260.0$ mT, was then subtracted from the on-resonance spectrum. These measurements were performed using frequency modulation $f_{mod} = 100$ kHz, frequency deviation $f_{dev} = 0.4$ MHz, which allowed for integration of the detected signal over a

bandwidth equivalent to ≈ 0.014 mT magnetic field modulation in a typical EPR experiment, time constant of the lock-in amplifier $\tau = 20$ ms, total frequency sweep width of 48 MHz and $N = 1201$ data points. The BDPA grain was then measured with the permanent magnet using the same experimental parameters; however, only $N = 534$ data points were recorded. In this case, the off-resonance spectrum required for the baseline correction was obtained by moving the permanent magnet 2 cm away from the EPRoC. In this configuration, the EPRoC was no longer under the permanent magnet, and thus outside the region of the B_0 field. The 2D mapping procedure was performed by translating the magnet along the x - and y -direction using steps of 50 and 100 μm , respectively, keeping the distance between the EPRoC and the permanent magnet fixed at 300 μm . The measurements on the thin film a-Si sample were recorded as the average of 20 spectra for each distance using frequency modulation $f_{mod} = 100$ kHz, frequency deviation $f_{dev} = 5.2$ MHz, time constant of the lock-in amplifier $\tau = 20$ ms, total frequency sweep width 137.2 MHz and $N = 502$ data points. The magnetic field strength along the vertical direction was determined by moving the position of the magnet in steps of 50 μm along the z -direction. The measurements of tempol in 80% glycerol and 20% water solution were recorded as the average of 30 spectra using frequency modulation $f_{mod} = 100$ kHz, frequency deviation $f_{dev} = 1.3$ MHz, time constant of the lock-in amplifier $\tau = 20$ ms, total frequency sweep 204 MHz and $N = 511$ data points. The four tempol solutions with 63, 31.7, 15.8 and 7.9 mM dissolved in 80% glycerol and 20% water were also measured using a Magnetech MS5000 spectrometer with 100 kHz modulation frequency, $B_1 = 3.8$ μT , and 40 s of sweep time. The maximum value of $B_1 = 20$ μT at the sample position 50 μm above the EPRoC surface was estimated from the Biot-Savart law according to the procedure showed in.⁴⁵ The tempol spectra measured with the EPRoC and permanent magnet were filtered using a second order Savitzky–Golay filter with a window of 6 data points in order to prevent line shape broadening beyond 5%. For controlling the experiments and the data acquisition with the EPRoC and the permanent magnet, a customized version of the software package Qudi was used.⁴⁶ Simulations were performed using the pepper function of the EasySpin library in MATLAB (Mathworks) for simulating power-averaged spectra of the thin-film a-Si sample to obtain a reference spectrum for the calculation of the line width broadening induced by the permanent magnet.⁴⁷ The a-Si spectrum was simulated using the values of the g -factor and line width (0.38 mT) at 7.3 GHz reported in literature.^{13,45} The baseline correction for the double integration procedure was performed using the backcor⁴⁸ function in MATLAB for a first order polynomial utilizing an asymmetric truncated quadratic function and values of the threshold that were chosen to avoid signal distortions. The least-squares fitting routine using the esfit function of the EasySpin library in MATLAB was performed to determine the line width and the correlation time of the 31.7 mM tempol solution in 80% Glycerol and 20% water.⁴⁷ Due to the high viscosity of the solution, the chili function for simulating the slow-motion regime was used in the fitting procedure. Moreover, a field offset parameter to compensate for the uncertainty of the magnetic field due to the positioning of the permanent magnet with respect to the EPRoC was included in the least-squares analysis.

RESULTS AND DISCUSSION

2D Magnetic Field Mapping. The EPRoC has been used as a sensor to determine the intensity and homogeneity of the magnetic field B_0 of the permanent magnet by recording the EPR spectra of a point-like BDPA sample placed on a single coil of the 7 GHz array. All measurements were performed by sweeping the frequency of the EPRoC. The first measurements were carried out using the EPRoC and the conventional electromagnet to characterize the properties of the EPRoC spectrum of BDPA. The spectrum shown in Figure 3a was recorded using the electromagnet, and the distance between the two minima ($\Delta f_{\text{min}} \approx 5.3$ MHz) of the FM-detected signal




Figure 3. (a) FM-detected EPR spectrum of BDPA measured using the 7 GHz EPRoC array in an electromagnet with high homogeneity (≈ 40 ppm). The distance $\Delta\tilde{f}_{mm} = 5.5$ MHz between the two minima is used as a reference value. (b) EPR spectrum of BDPA measured using the 7 GHz EPRoC array and the permanent magnet. The distance $\Delta f_{mm} = 15.5$ MHz between the two minima is measured for each EPR spectrum during the mapping procedure. It should be noticed that in Figure 2a the resonance occurred at $B_0 = 255.7$ mT, which is ≈ 1 mT higher than the average magnetic field in the homogeneous region of the permanent magnet. This discrepancy is due to the magnetic field offset occurring when positioning the EPRoC between the poles of the electromagnet. (c) 2D mapping of the intensity of the magnetic field of the permanent magnet. The map is obtained via calculation of the resonance position of the EPR spectrum at each spatial position. The pixel size is $50 \mu\text{m} \times 100 \mu\text{m}$ (x, y), and the total mapped area is $5 \text{ mm} \times 1.15 \text{ mm}$. (d) 2D map of the broadening (in ppm) of the EPR spectral line induced by the inhomogeneity of the magnetic field of permanent magnet. The orange dashed rectangle containing 14 octagonal shapes depicts (to scale) the EPRoC sensors and where the red dot represents the BDPA grain used to perform the mapping procedure.

was used as reference value for the line width of the spectrum which was obtained from the total frequency deviation. Figure 3b shows the result of the measurements obtained using the permanent magnet. The position of the resonance peak and the distance between the two minima ($\Delta f_{mm} \approx 12.6$ MHz) were used to determine the value of the field ($B_0 = 254.95$ mT) and its homogeneity by comparing with the reported BDPA g value, 2.0026, and the increase in line width from the recorded reference spectrum shown in Figure 3a, respectively.⁴⁹ Figure 3c,d show the 2D map of the intensity and the homogeneity of the magnetic field at the position of the EPRoC sensor above the permanent magnet, respectively. The mapping was performed by translating the permanent magnet with respect to the BDPA grain placed on one of the EPRoC sensors while measuring the EPR spectrum at each position with a step size of 0.1 mm . The distance between the EPRoC and the magnet was $d = 300 \mu\text{m}$ and was not varied throughout the mapping procedure. The intensity of the B_0 field was determined at each spatial position from the resonance position of the EPR spectrum of the BDPA. The value of the magnetic field over the entire mapped area of the permanent magnet lies within the interval $B_0 \in [253.22, 256.45]$, with the weighted average value $\bar{B}_0 = 254.69$ mT. The distance between the two minima of each spectrum Δf_{mm} was used to determine the signal broadening, ΔB_{ppm} in ppm, induced by the inhomogeneity of the field using the following relation

$$\Delta B_{\text{ppm}} = \frac{\Delta f_{mm,i} - \Delta\tilde{f}_{mm}}{f_{0,1} + f_{0,2}} \quad (2)$$

where $\Delta f_{mm,i}$ is the distance between the two minima of the FM-signal measured with the permanent magnet at the i th-position, $\Delta\tilde{f}_{mm}$ is the reference value of the distance between the two minima of the FM-signal measured using the electromagnet, and $f_{0,1}$ and $f_{0,2}$ are the corresponding frequency values of the two minima in the reference spectrum, see Figure 3a. The full derivation of the eq 2 can be found in the “Broadening derivation” Section in the Supporting Information. The results of these measurements demonstrate that the highest homogeneity is ≈ 700 ppm, as shown in Figure 3d, blue, and it stretches over about 40% of the investigated region. This value is comparable with the value obtained from the Hall probe measurements. It should be noted that the absolute field intensity using the EPRoC sensor directly as a Hall probe like device, shown in Figure 3c, where the 2D map directly describes the magnetic field, provides a sufficient measurement of homogeneity. However, the line width of the selected EPR sample provides additional information regarding the effects of homogeneity on the accuracy (in ppm) of the intended sensing application as it directly indicates the limitations of measuring the EPR signal with the EPRoC device.

Magnetic Field Mapping along the Vertical Direction.

We also evaluated the magnetic field intensity and homogeneity along the vertical direction of the permanent magnet. In




Figure 4. (a) Dependence of the B_0 intensity as a function of the distance between the EPRoC array and the permanent magnet. The data were obtained via calculation of the db resonance position of the a-Si sample placed on the entire VCO array. (b) Dependence of the dangling bonds signal broadening measured in ppm as a function of distance between the EPRoC array and the permanent magnet. The errors are indicated at each data point.

this case the FM signal of the neutral dangling bonds (db), which are homogeneously dispersed in a 15 μm thin film of amorphous silicon (a-Si), was measured. Due to its dimensions, the a-Si sample covered the entire VCO array to achieve uniformly distributed db spins on each coil of the VCO. In this way, it was possible to determine the line width broadening induced by the magnetic field inhomogeneity when all coils of the array are active. Based on the results of the 2D mapping, the x - and y -positions of the EPRoC with respect to the permanent magnet were selected such that the EPRoC would be placed in the region of highest homogeneity thereby minimizing signal broadening. The distance between the EPRoC and the magnet was increased in steps of 50 μm , and the EPR spectrum was recorded at each position. The intensity of B_0 and the broadening of the EPR signal were determined via calculation of the resonance position using the characteristics of the a-Si db signal at 7.3 GHz ($g = 2.0055$, line width $\Delta B_{\text{pp}} = 0.38$ mT). Figure 4a shows the dependence of field intensity on the distance between the EPRoC and the permanent magnet. In Figure 4b the broadening (in ppm) of the EPR line width induced by the inhomogeneity of the magnetic field for each position of the permanent magnet with respect to the EPRoC is shown. Spectral simulations and the least-squares fitting routine employed to characterize the obtained data are reported in the **Material and Methods** section. The distance between the minima of each FM-detected signal was determined from the fit and the broadening of the EPR line width in ppm was calculated using eq 2 and plotted in Figure 4b. The errors were determined from the errors provided by the fitting routine of the esfit function.⁴⁷ An example of FM-detected signal of the a-Si sample can be found in the “a-Si FM-detected EPR signal” section in the **Supporting Information**.

Spin Counting and Molecular Tumbling with EPRoC.

The information about the homogeneity and the intensity of B_0 obtained after the mapping procedure was used to calibrate the position of the EPRoC with respect to the permanent magnet in the region of highest homogeneity to perform proof of principle EPR experiments. Tempol solutions of different concentrations (63, 31.7, 15.8, and 7.9 mM) dissolved in a solution of 20% water and 80% glycerol were measured to establish a calibration procedure for quantitative measurements

using the benchtop Magnetech MS5000 spectrometer and the EPRoC in combination with the permanent magnet. For these measurements, the distance between the EPRoC and the permanent magnet was kept constant at 300 μm . The signal intensity of the data measured with the Magnetech MS5000 spectrometer was obtained by numerical double integration of each recorded spectrum while polynomial baseline correction was performed before each integration. Since the spectra measured with EPRoC are dispersion-like signals, the FM-signals were Hilbert transformed to absorption-like signals using the Kramers–Kronig relations.⁵⁰ It should be noted that Kramers–Kronig relations can be applied only for signals that are not power saturated.⁵⁰ Accordingly, the measurements with the EPRoC have been performed for $B_1 = 20$ μT which was confirmed to still be in the linear, unsaturated regime as determined from the power–saturation curve of the tempol solutions using the Magnetech MS5000. Subsequently, the same double integration and baseline correction procedures were performed to obtain the signal intensity. The results of the measurements are shown in Figure 5. The signal intensity normalized to the maximum value obtained for the highest concentration is plotted as a function of the sample concentration. A linear fit has been performed for each data set and good agreement was found between the angular coefficients $m_{\text{MS500}} = 0.0159(3)$ and $m_{\text{EPRoC}} = 0.0158(2)$, for the MS5000 and EPRoC, respectively.

After determining the linear dependence of the EPR signal on concentration, the spectra of the 31.7 mM tempol dissolved in 80% glycerol and 20% water solution and placed in the flat capillaries were analyzed in further detail. The typical three-line pattern resulting from the hyperfine interaction of the electron spin with the nitrogen nuclear spin,⁵¹ measured using both the Magnetech MS5000 and the EPRoC and permanent magnet, is shown in Figure 6. It should be noted that the sample was used when measuring with both spectrometers to facilitate comparative data evaluation without consideration of sample environment variations. The EPRoC FM-detected signal was measured by sweeping the frequency of the EPRoC, and subsequently Hilbert transformed to an absorption signal. The data were fitted using the esfit function provided by the EasySpin⁴⁷ library employing the chili function for the slow-motion regime, due to the high viscosity of glycerol as




Figure 5. EPR signal intensity obtained by second integration of the data measured using the MS5000 (blue dots) and using the EPRoC with the permanent (orange dots) plotted as a function of tempol concentration. The data have been normalized to the maximum value obtained for the highest sample concentration measured (63 mM). The solid lines show the best fit to the data assuming a linear relation. Error bars are within the size of the data points.

explained in the Material and Methods section. For the simulation, a rhombic g factor ($g = [2.0092 \ 2.0061 \ 2.0059]$) and hyperfine tensor ($A = [0.55 \ 0.63 \ 0.359]$ mT) were used, and are consistent with literature reports.^{52,53} The line width (ΔB_{pp}) and tumbling correlation time (τ_R) obtained from the fit were compared to results in the literature (Table 1).

DISCUSSION

The characterization of the permanent magnet, presented in the previous sections, demonstrates the applicability of the EPRoC as a sensor for imaging static magnetic fields. The 2D maps of the magnetic field intensity and the signal broadening shown in Figure 3c,d, respectively were obtained with a spatial resolution of $100 \mu\text{m} \times 50 \mu\text{m}$ along the x - and y -directions in a range of $1.2 \text{ mm} \times 5 \text{ mm}$. The spatial resolution was mainly determined by the size of the BDPA grain ($\approx 200 \mu\text{m}$ diameter) placed on top of one array detector. Hence, higher spatial resolution in the micrometer range may be obtained using smaller samples. The results in Figure 3c allow for the calculation of the average value of $\bar{B}_0 = 254.69 \text{ mT}$ over the investigated region of the permanent magnet. In Figure 3d the broadening obtained from the observed BDPA line width (eq 2) is shown. In the region of highest homogeneity, the line width broadening is $\approx 700 \text{ ppm}$. This region extends for $\approx 5 \text{ mm}$ along the x -direction and $\approx 0.6 \text{ mm}$ along the y -direction. Since the diameter of each coil of the VCO is $200 \mu\text{m}$, it is possible to place the whole 2×7 array in this region with highest homogeneity in order to utilize all 14 coils for EPR measurements (Figure 3d). The results in ppm obtained using eq 2 are largely dependent on the choice of the paramagnetic probe (BDPA). For probes that exhibit very small line widths, large line broadening values are obtained. Conversely, for probes whose line width is greater than the magnetic field inhomogeneity, the line broadening value tends to zero. Therefore, to characterize homogeneous magnetic fields, it is necessary to select samples with a very small line width. This approach allows for direct determination of the quality of




Figure 6. (a) EPR signal of 31.7 mM tempol dissolved in 80% glycerol and 20% water. The black trace represents the experimental data measured using the benchtop EPR spectrometer (Magnetech MS5000) and the red trace is the result of the fit to the data. (b) EPR signal of 31.7 mM tempol dissolved in 80% glycerol and 20% water measured using the EPRoC and permanent magnet. The absorption signal has been obtained by Hilbert transformation of the FM-detected signal via Kramers–Kronig relations. The black and red traces represent the experimental data and the fit, respectively. The spectrum has been filtered using a 2nd order Savitzky–Golay filter with a window chosen to be small enough to ensure a line broadening $<5\%$. Since the same sample was measured using the two different EPR experimental configurations, the line width discrepancy may therefore be attributed to the inhomogeneity of the magnetic field of the permanent magnet.

Table 1. Results of the Fitting Procedure Performed on the Data Measured Using the Magnetech MS5000 and the EPRoC for the Line Width (ΔB_{pp}) and the Tumbling Correlation Time (τ_R) of the 31.7 mM Tempol in 80% Glycerol and 20% Water Solution. Values from the Literature are Shown for Comparison

	MS5000	EPRoC	literature ³⁸
ΔB_{pp}	0.19(3) mT	0.59(4) mT	0.16 mT
τ_R	0.34(2) ns	0.56(5) ns	0.39 ns

magnets used in EPRoC spectrometers and to evaluate their potential for EPR spectroscopy applications. It should be noted that when characterizing more homogeneous permanent

magnets using samples with narrow line widths, it is essential to ensure that the magnetic field homogeneity of the electromagnet is high enough to prevent line width broadening of the reference sample. The height dependence of the magnetic field was determined with an a-Si sample covering the whole VCO array. The results in Figure 4a show that the field strength increases along the vertical direction perpendicular to the x – y plane of the permanent magnet. However, the signal broadening (Figure 4b) up to a distance of $\approx 600\ \mu\text{m}$ from the magnet surface is $\approx 150\ \text{ppm}$. Thus, the homogeneity of the permanent magnet is sufficient to measure samples with line widths slightly larger than 0.4 mT with the magnet used here without introducing signal broadening. These results, combined with the mapping procedure performed using the BDPA sample, demonstrate the possibility to perform 3D magnetic field imaging using either a single coil or all 14 coils of the VCO array. In this way, the region of best homogeneity of the magnetic field can be accessed to perform EPR measurements. The EPRoC and the permanent magnet have been used to measure tempol with different spin concentrations (63, 31.7, 15.8, and 7.9 mM) dissolved in 80% glycerol and 20% water solutions to establish a calibration procedure for spin counting. The findings shown in Figure 5 indicate a clear linear relationship between the signal intensity acquired using the benchtop MS5000 spectrometer and the EPRoC in combination with the permanent magnet, thereby demonstrating the viability of quantifying the number of spins in other samples, when placed in a flat capillary or other uniform sample geometry, by computing the signal intensity. The 31.7 mM tempol solution was then analyzed in further detail to evaluate the possibility of determining the tumbling correlation time of the tempol molecules. As it has been already demonstrated in other works,³⁸ the high viscosity of glycerol decreases the tumbling rate of the nitroxide and this effect results in different intensities of the three spectral lines. This effect is well depicted in the differing peak intensities observed for the measurements shown in Figure 6a,b, and it is further described by the results of the fit reported in Table 1. The values of the line width $\Delta B_{\text{pp}} = 0.19(3)\ \text{mT}$ and correlation time $\tau_c = 0.34(2)\ \text{ns}$ measured with the benchtop MS5000 spectrometer are in good agreement with the findings in ref 38. The tumbling correlation time measured with the EPRoC is about a factor of 1.6 larger than the value obtained with the MS5000. This discrepancy is due to the signal broadening introduced by the inhomogeneity of the magnetic field of the permanent magnet. Indeed, the line width ΔB_{pp} measured with the EPRoC is about a factor of 3 larger than the value obtained with the MS5000. As described in,⁵⁴ the line width variation due to broadening effects is particularly relevant for correlation time $\tau_c < 1\ \text{ns}$. Despite the discrepancies in the results, the EPRoC and the permanent magnet used in this work exhibit potential for accurate measurements of the tumbling correlation time of spin label nitroxides in the slow-motion regime ($\tau_c > 1\ \text{ns}$) due to the greater value of line widths (typically $\Delta B_{\text{pp}} > 0.4\ \text{mT}$).⁵¹

CONCLUSIONS AND OUTLOOK

We have successfully demonstrated a portable EPR spectrometer based on a small single-sided permanent magnet, combined with a 7 GHz array EPRoC. The EPRoC sensor was used to determine the spatial distribution of the strength and homogeneity of the magnetic field B_0 of the permanent magnet, which is shown to be well-suited for measuring

undistorted spectra with a line width larger than 0.4 mT. The combined system was used to determine the molecular tumbling correlation time of tempol in glycerol, showing a discrepancy of a factor 1.6 with respect to measurements with a conventional EPR spectrometer and to what has previously been reported in the literature. This is attributed to line width broadening induced by the inhomogeneity of the magnetic field. However, it was possible to demonstrate that the broadening of the EPR line width does not represent a limiting factor for the quantitative determination of spin concentration, the inhomogeneous line broadening would only reduce the signal-to-noise ratio (S/N). To realize higher magnetic field homogeneity in future, specifically with respect to S/N in case of narrow lines, the design can be further refined by using an ablation laser to directly inscribe the corrective pattern onto the surface of the magnet segments. This avoids the use of the μ -strip, which is hard to be accurately align and making the mechanical assembly and the lasing process more precise and less susceptible to errors. Overall, the EPRoC represents a robust solution for magnetic field imaging application.²⁶ The EPRoC used in this work has a bandwidth of $\approx 1\ \text{GHz}$, which allows the mapping of magnetic fields over a range of 34 mT. New designs of the EPRoC device with a larger bandwidth would allow mapping of the magnetic field in a much broader field intensity range. Moreover, we have demonstrated the applicability of the EPRoC device and the permanent magnet for spin counting applications, and for molecular tumbling determination of nitroxides in the slow-motion regime. Further improvements of the permanent magnet homogeneity will ensure precise determinations of the molecular correlation time in the fast-motion regime and at the isotropic limit as well.^{51,54} Moreover, the compact size and single-sided design of the permanent magnet allow, in combination with the EPRoC, the develop of a dipstick EPR sensor which can be immersed directly into the sample environment.⁵⁵ Overall, the EPRoC and the permanent magnet represent a promising solution for the development of a new generation of versatile and compact EPR spectrometers, which could be driven by a battery for in situ, in vivo and operando spectroscopic applications in the field of energy materials, healthcare, food quality control and catalysis to determine surface spin concentration with great potential for industrial applications.

ASSOCIATED CONTENT

Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acssensors.4c00788>.

B_0 mapping. A derivation of the broadening parameter and the recorded FM spectrum of the a-Si sample are included (PDF)

AUTHOR INFORMATION

Corresponding Author

Joseph E. McPeak – Helmholtz-Zentrum Berlin für Materialien und Energie GmbH, 14109 Berlin, Germany; orcid.org/0000-0001-8677-6405; Email: jm@chem.ku.dk

Authors

Michele Segantini – Helmholtz-Zentrum Berlin für Materialien und Energie GmbH, 14109 Berlin, Germany

Gianluca Marcozzi – Helmholtz-Zentrum Berlin für Materialien und Energie GmbH, 14109 Berlin, Germany
Tarek Elrifai – Institute of Smart Sensors, Universität Stuttgart, 70569 Stuttgart, Germany
Ekaterina Shabratova – Helmholtz-Zentrum Berlin für Materialien und Energie GmbH, 14109 Berlin, Germany; Leibniz-Institut für Höchstfrequenztechnik, Ferdinand-Braun-Institut gGmbH, 12489 Berlin, Germany
Katja Höflich – Leibniz-Institut für Höchstfrequenztechnik, Ferdinand-Braun-Institut gGmbH, 12489 Berlin, Germany
Mihaela Deaconea – Bruker BioSpin GmbH, 76275 Ettlingen, Germany
Volker Niemann – Bruker BioSpin GmbH, 76275 Ettlingen, Germany
Rainer Pietig – Bruker BioSpin GmbH, 76275 Ettlingen, Germany
Jens Anders – Institute of Smart Sensors, Universität Stuttgart, 70569 Stuttgart, Germany; Center for Integrated Quantum Science and Technology, 70569 Stuttgart and Ulm, Germany
Boris Naydenov – Helmholtz-Zentrum Berlin für Materialien und Energie GmbH, 14109 Berlin, Germany
Klaus Lips – Helmholtz-Zentrum Berlin für Materialien und Energie GmbH, 14109 Berlin, Germany; Berlin Joint EPR Laboratory, Fachbereich Physik, Freie Universität Berlin, 14195 Berlin, Germany; orcid.org/0000-0003-4691-7239

Complete contact information is available at:

<https://pubs.acs.org/10.1021/acssensors.4c00788>

Author Contributions

M.S., G.M., J.M., B.N., K.L. defined the goals of the research and designed the experiments. MS performed all EPR experiments. M.S. performed the EPR simulations and data analysis. M.S. and G.M. customized Qudi to perform all the measurements with the EPRoC. T.E. designed the 7 GHz array EPRoC and the PCB. V.N. and R.P. designed the permanent magnet. M.D. assembled the permanent magnet. E.S. and K.H. aided in the design, fabrication and characterization processes of the permanent magnet. M.S., J.M., G.M., B.N., and K.L. evaluated the results of the experiments and wrote the manuscript.

Funding

This work has been supported by the Bundesministerium für Bildung und Forschung under contract number 01186916/1 (EPRoC), the Deutsche Forschungsgemeinschaft (DFG) for the project “THz EPR-on-a-Chip for Enhanced Spin Sensitivity” within the priority program Interest (SPP2314) and by the HEMF (Helmholtz Energy Materials Foundry) infrastructure funded by the Helmholtz association (HGF). B.N. acknowledges the financial support from the DFG (project numbers 410866378 and 410866565).

Notes

The authors declare no competing financial interest.

ACKNOWLEDGMENTS

We are grateful to Dr. Jannik Möser (HZB), Silvio Künstner (HZB), and Prof. Dr. Jan Behrends (Freie Universität Berlin, FUB and HZB) for experimental support and for helpful discussions. We are thankful to Martin Muske (HZB) for fabrication of the amorphous silicon samples. We are thankful to Daria Astapovych for the helpful discussions and support in the magnet fabrication. We acknowledge Mohamed Hassan

(UST) and Hadi Lotfi (UST) for their contributions to the EPRoC design and fabrication.

REFERENCES

- (1) Di Meo, S.; Reed, T. T.; Venditti, P.; Victor, V. M. Role of ROS and RNS Sources in Physiological and Pathological Conditions. *Oxid. Med. Cell. Longevity* **2016**, 2016, 1245049.
- (2) Suzen, S.; Gurer-Orhan, H.; Saso, L. Detection of Reactive Oxygen and Nitrogen Species by Electron Paramagnetic Resonance (EPR) Technique. *Molecules* **2017**, 22 (1), 181.
- (3) Schaner, P. E.; Pettus, J. R.; Flood, A. B.; Williams, B. B.; Jarvis, L. A.; Chen, E. Y.; Pastel, D. A.; Zuurbier, R. A.; diFlorio-Alexander, R. M.; Swartz, H. M.; Kuppusamy, P. OxyChip Implantation and Subsequent Electron Paramagnetic Resonance Oximetry in Human Tumors Is Safe and Feasible: First Experience in 24 Patients. *Front. Oncol.* **2020**, 10, 572060.
- (4) Demidenko, E.; Kmiec, M. M.; Kuppusamy, P. Estimation of pO₂ distribution in EPR oximetry. *J. Magn. Reson.* **2021**, 328, 106992.
- (5) Papadimitriou, V.; Sotiropoulos, T. G.; Xenakis, A.; Sofikiti, N.; Stavyiannoudaki, V.; Chaniotakis, N. A. Oxidative Stability and Radical Scavenging Activity of Extra Virgin Olive Oils: An Electron Paramagnetic Resonance Spectroscopy Study. *Anal. Chim. Acta* **2006**, 573–574, 453–458.
- (6) Fadda, A.; Molinu, M. G.; Deiana, P.; Sanna, D. Electron Paramagnetic Resonance Spin Trapping of Sunflower and Olive Oils Subjected to Thermal Treatment: Optimization of Experimental and Fitting Parameters. *ACS Food Sci. Technol.* **2021**, 1 (7), 1294–1303.
- (7) Ricca, M.; Foderà, V.; Vetri, V.; Buscarino, G.; Montalbano, M.; Leone, M. Oxidation Processes in Sicilian Olive Oils Investigated by a Combination of Optical and EPR Spectroscopy. *J. Food Sci.* **2012**, 77 (10), C1084–C1089.
- (8) Cui, L.; Lahti, P. M.; Decker, E. A. Evaluating Electron Paramagnetic Resonance (EPR) to Measure Lipid Oxidation Lag Phase for Shelf-Life Determination of Oils. *J. Am. Oil Chem. Soc.* **2017**, 94 (1), 89–97.
- (9) Kunz, T.; Müller, C.; Mato-Gonzales, D.; Methner, F.-J. The Influence of Unmalted Barley on the Oxidative Stability of Wort and Beer. *J. Inst. Brew.* **2012**, 118 (1), 32–39.
- (10) Kocherginsky, N. M.; Kostetskiy, Y. Yu.; Smirnov, A. I. Antioxidant Pool in Beer and Kinetics of EPR Spin-Trapping. *J. Agric. Food Chem.* **2005**, 53 (17), 6870–6876.
- (11) Porcu, M. C.; Fadda, A.; Sanna, D. Lag Time Determinations in Beer Samples. Influence of Alcohol and PBN Concentration in EPR Spin Trapping Experiments. *Oxygen* **2022**, 2 (4), 605–615.
- (12) Kunz, T.; Strähmel, A.; Cortés, N.; Kroh, L. W.; Methner, F.-J. Influence of Intermediate Maillard Reaction Products with Enediol Structure on the Oxidative Stability of Beverages. *J. Am. Soc. Brew. Chem.* **2013**, 71 (3), 114–123.
- (13) Fehr, M.; Schnegg, A.; Rech, B.; Lips, K.; Astakhov, O.; Finger, F.; Freysoldt, C.; Bittl, R.; Teutloff, C. Dangling Bonds in Amorphous Silicon Investigated by Multifrequency EPR. *J. Non-Cryst. Solids* **2012**, 358 (17), 2067–2070.
- (14) Cochrane, C. J.; Lenahan, P. M.; Lelis, A. J. Identification of a Silicon Vacancy as an Important Defect in 4H SiC Metal Oxide Semiconducting Field Effect Transistor Using Spin Dependent Recombination. *Appl. Phys. Lett.* **2012**, 100 (2), 023509.
- (15) Segantini, M.; Marcozzi, G.; Djekic, D.; Chu, A.; Amkreutz, D.; Trinh, C. T.; Neubert, S.; Stannowski, B.; Jacob, K.; Rudolph, I.; McPeak, J. E.; Anders, J.; Naydenov, B.; Lips, K. Electrically Detected Magnetic Resonance on a Chip (EDMRoC) for Analysis of Thin-Film Silicon Photovoltaics. *Magnetochemistry* **2023**, 9 (7), 183.
- (16) Lawton, J. S.; Jones, A.; Zawodzinski, T. Concentration Dependence of VO₂⁺ Crossover of Nafion for Vanadium Redox Flow Batteries. *J. Electrochem. Soc.* **2013**, 160 (4), A697–A702.
- (17) Chace, W. S.; Tiano, S. M.; Arruda, T. M.; Lawton, J. S. Effects of State of Charge on the Physical Characteristics of V(IV)/V(V) Electrolytes and Membrane for the All Vanadium Flow Battery. *Batteries* **2020**, 6 (4), 49.

- (18) Wandt, J.; Marino, C.; Gasteiger, H. A.; Jakes, P.; Eichel, R.-A.; Granwehr, J. Operando Electron Paramagnetic Resonance Spectroscopy – Formation of Mossy Lithium on Lithium Anodes during Charge–Discharge Cycling. *Energy Environ. Sci.* **2015**, *8* (4), 1358–1367.
- (19) Niemöller, A.; Jakes, P.; Eurich, S.; Paulus, A.; Kungl, H.; Eichel, R.-A.; Granwehr, J. Monitoring Local Redox Processes in LiNi_{0.5}Mn_{1.5}O₄ Battery Cathode Material by in Operando EPR Spectroscopy. *J. Chem. Phys.* **2018**, *148* (1), 014705.
- (20) Weckhuysen, B. M. Snapshots of a Working Catalyst: Possibilities and Limitations of in Situ Spectroscopy in the Field of Heterogeneous Catalysis. *Chem. Commun.* **2002**, No. 2, 97–110.
- (21) Brückner, A. Monitoring Transition Metal Ions (TMI) in Oxide Catalysts during (Re)Action: The Power of Operando EPR. *Phys. Chem. Chem. Phys.* **2003**, *5* (20), 4461–4472.
- (22) Grauke, R.; Schepper, R.; Rabeah, J.; Schoch, R.; Bentrup, U.; Bauer, M.; Brückner, A. Impact of Al Activators on Structure and Catalytic Performance of Cr Catalysts in Homogeneous Ethylene Oligomerization – A Multitechnique in Situ/Operando Study. *ChemCatChem* **2020**, *12* (4), 1025–1035.
- (23) Bonke, S. A.; Risse, T.; Schnegg, A.; Brückner, A. In Situ Electron Paramagnetic Resonance Spectroscopy for Catalysis. *Nat. Rev. Methods Primers* **2021**, *1* (1), 33.
- (24) Geng, F.; Yang, Q.; Li, C.; Shen, M.; Chen, Q.; Hu, B. Mapping the Distribution and the Microstructural Dimensions of Metallic Lithium Deposits in an Anode-Free Battery by In Situ EPR Imaging. *Chem. Mater.* **2021**, *33* (21), 8223–8234.
- (25) Sathiyar, M.; Leriche, J.-B.; Salager, E.; Gourier, D.; Tarascon, J.-M.; Vezin, H. Electron Paramagnetic Resonance Imaging for Real-Time Monitoring of Li-Ion Batteries. *Nat. Commun.* **2015**, *6* (1), 6276.
- (26) Tseytlin, O.; Bobko, A. A.; Tseytlin, M. Rapid Scan EPR Imaging as a Tool for Magnetic Field Mapping. *Appl. Magn. Reson.* **2020**, *51* (9–10), 1117–1124.
- (27) Haas, S.; Gordijn, A.; Stiebig, H. High Speed Laser Processing for Monolithical Series Connection of Silicon Thin-Film Modules. *Prog. Photovolt. Res. Appl.* **2008**, *16* (3), 195–203.
- (28) Anders, J.; Lips, K. MR to Go. *J. Magn. Reson.* **2019**, *306*, 118–123.
- (29) Schlecker, B.; Hoffmann, A.; Chu, A.; Ortmanns, M.; Lips, K.; Anders, J. Towards Low-Cost, High-Sensitivity Point-of-Care Diagnostics Using VCO-Based ESR-on-a-Chip Detectors. *IEEE Sens. J.* **2019**, *19* (20), 8995–9003.
- (30) Hassan, M. A.; Kern, M.; Chu, A.; Kalra, G.; Shabrato, E.; Tsarapkin, A.; MacKinnon, N.; Lips, K.; Teutloff, C.; Bittl, R.; Korvink, J. G.; Anders, J. Towards Single-Cell Pulsed EPR Using VCO-Based EPR-on-a-Chip Detectors. *Frequenz* **2022**, *76* (11–12), 699–717.
- (31) Chu, A.; Schlecker, B.; Kern, M.; Goodsell, J. L.; Angerhofer, A.; Lips, K.; Anders, J. On the Modeling of Amplitude-Sensitive Electron Spin Resonance (ESR) Detection Using Voltage-Controlled Oscillator (VCO)-Based ESR-on-a-Chip Detectors. *Magn. Reson.* **2021**, *2* (2), 699–713.
- (32) Künstner, S.; Chu, A.; Dinse, K.-P.; Schnegg, A.; McPeak, J. E.; Naydenov, B.; Anders, J.; Lips, K. Rapid-Scan Electron Paramagnetic Resonance Using an EPR-on-a-Chip Sensor. *Magn. Reson.* **2021**, *2* (2), 673–687.
- (33) Chu, A.; Schlecker, B.; Handwerker, J.; Kunstner, S.; Ortmanns, M.; Lips, K.; Anders, J. VCO-Based ESR-on-a-Chip as a Tool for Low-Cost, High-Sensitivity Food Quality Control. In *2017 IEEE Biomedical Circuits and Systems Conference (BioCAS)*, 2017, pp 1–4.
- (34) Handwerker, J.; Schlecker, B.; Wachter, U.; Radermacher, P.; Ortmanns, M.; Anders, J. 28.2 A 14 GHz Battery-Operated Point-of-Care ESR Spectrometer Based on a 0.13 μ m CMOS ASIC. In *2016 IEEE International Solid-State Circuits Conference (ISSCC)*, 2016, pp 476–477.
- (35) Hassan, M. A.; Elrifai, T.; Sakr, A.; Kern, M.; Lips, K.; Anders, J. A 14-Channel 7 GHz VCO-Based EPR-on-a-Chip Sensor with Rapid Scan Capabilities. In *2021 IEEE Sensors*, 2021, pp 1–4.
- (36) Elrifai, T.; Sakr, A.; Lotfi, H.; Hassan, M. A.; Lips, K.; Anders, J. A 7 GHz VCO-Based EPR Spectrometer Incorporating a UWB Data Link. In *2022 20th IEEE Interregional NEWCAS Conference (NEWCAS)*, 2022, pp 280–284.
- (37) Chu, A.; Schlecker, B.; Lips, K.; Ortmanns, M.; Anders, J. An 8-Channel 13 GHz ESR-on-a-Chip Injection-Locked Vco-Array Achieving 200 μ M-Concentration Sensitivity. In *2018 IEEE International Solid - State Circuits Conference—(ISSCC)*, 2018, pp 354–356.
- (38) Clark, A.; Sedhom, J.; Elajaili, H.; Eaton, G. R.; Eaton, S. S. Dependence of Electron Paramagnetic Resonance Spectral Lineshapes on Molecular Tumbling: Nitroxide Radical in Water:Glycerol Mixtures. *Concepts Magn. Reson., Part A* **2016**, *45A* (5), No. e21423.
- (39) Poncelet, M.; Driesschaert, B. A 13C-Labeled Triarylmethyl Radical as an EPR Spin Probe Highly Sensitive to Molecular Tumbling. *Angew. Chem., Int. Ed.* **2020**, *59* (38), 16451–16454.
- (40) Suksmono, A. B.; Danudirdjo, D.; Setiawan, A. D.; Rahmawati, D.; Prastio, R. P. A Magnetic Field Camera for Real-Time Subsurface Imaging Applications. *Appl. Sci.* **2021**, *11* (8), 3302.
- (41) Leonard, F.; Godin, B. 3D In Vitro Model for Breast Cancer Research Using Magnetic Levitation and Bioprinting Method. In *Breast Cancer: Methods and Protocols*; Cao, J., Ed.; *Methods in Molecular Biology*; Springer: New York, NY, 2016; pp 239–251.
- (42) Zgadzai, O.; Twig, Y.; Wolfson, H.; Ahmad, R.; Kuppusamy, P.; Blank, A. Electron-Spin-Resonance Dipstick. *Anal. Chem.* **2018**, *90* (13), 7830–7836.
- (43) Switala, L. E.; Black, B. E.; Mercovich, C. A.; Seshadri, A.; Hornak, J. P. An Electron Paramagnetic Resonance Mobile Universal Surface Explorer. *J. Magn. Reson.* **2017**, *285*, 18–25.
- (44) Schlecker, B.; Chu, A.; Handwerker, J.; Künstner, S.; Ortmanns, M.; Lips, K.; Anders, J. VCO-Based ESR-on-a-Chip as a Tool for Low-Cost, High-Sensitivity Point-of-Care Diagnostics. In *2017 IEEE Sensors*, 2017, pp 1–3.
- (45) Künstner, S.; McPeak, J. E.; Chu, A.; Kern, M.; Wick, M.; Dinse, K.-P.; Anders, J.; Naydenov, B.; Lips, K. Microwave Field Mapping for EPR on a Chip Experiments. *Sci. Adv.* **2024**, *10*, No. eado5467.
- (46) Binder, J. M.; Stark, A.; Tomek, N.; Scheuer, J.; Frank, F.; Jahnke, K. D.; Müller, C.; Schmitt, S.; Metsch, M. H.; Unden, T.; Gehring, T.; Huck, A.; Andersen, U. L.; Rogers, L. J.; Jelezko, F. Q. Qudi: A modular python suite for experiment control and data processing. *SoftwareX* **2017**, *6*, 85–90.
- (47) Stoll, S.; Schweiger, A. EasySpin a Comprehensive Software Package for Spectral Simulation and Analysis in EPR. *J. Magn. Reson.* **2006**, *178* (1), 42–55.
- (48) Vincent, M. Background Correction. <https://uk.mathworks.com/matlabcentral/fileexchange/27429-background-correction>.
- (49) Meyer, V.; Eaton, S. S.; Eaton, G. R. X-Band Electron Spin Relaxation Times for Four Aromatic Radicals in Fluid Solution and Comparison with Other Organic Radicals. *Appl. Magn. Reson.* **2014**, *45* (10), 993–1007.
- (50) Portis, A. M. Electronic Structure of F Centers: Saturation of the Electron Spin Resonance. *Phys. Rev.* **1953**, *91* (5), 1071–1078.
- (51) Bordignon, E. EPR Spectroscopy of Nitroxide Spin Probes. In *eMagRes*; John Wiley & Sons, Ltd, 2017; pp 235–254.
- (52) Symons, M. C. R.; Rockenbauer, A.; Györ, M.; Hankovszky, H. O.; Hideg, K. *E.S.R. Of the Conformation of 5- and 6-Membered Cyclic Nitroxide (Aminoxyl) Radicals*, 1988.
- (53) Biller, J. R.; Meyer, V.; Elajaili, H.; Rosen, G. M.; Kao, J. P. Y.; Eaton, S. S.; Eaton, G. R. Relaxation Times and Line Widths of Isotopically-Substituted Nitroxides in Aqueous Solution at X-Band. *J. Magn. Reson.* **2011**, *212* (2), 370–377.
- (54) Etienne, E.; Pierro, A.; Tamburrini, K. C.; Bonucci, A.; Mileo, E.; Martinho, M.; Belle, V. Guidelines for the Simulations of Nitroxide X-Band Cw EPR Spectra from Site-Directed Spin Labeling Experiments Using SimLabel. *Molecules* **2023**, *28* (3), 1348.
- (55) Künstner, S.; McPeak, J. E.; Chu, A.; Kern, M.; Dinse, K.-P.; Naydenov, B.; Fischer, P.; Anders, J.; Lips, K. Monitoring the State of Charge of Vanadium Redox Flow Batteries with an EPR-on-a-Chip Dipstick Sensor. *Phys. Chem. Chem. Phys.* **2024**, *26*, 17785–17795.